

Case Medical 510k Clearance's

510k #	Date	Description
K960738	May 1997	SteriTite container system for pre-vacuum and gravity displacement steam
K974638	December 1997	SteriTite container system for pre-vacuum steam.
K982151	February 1999	SteriTite Chemical Indicators.
K991023	January 2000	SteriTite container system for STERRAD 100 & MediTray Products containerized and wrapped.
K030853	March 2003	SteriTite container system for STERRAD 100S & MediTray Products containerized and wrapped.
K023614	April 2003	SteriTite container system for EO & MediTray products containerized and wrapped.
K022978	May 2003	SteriTite container system with FlashTite valve plate(s) for Steam, flash sterilization (IUSS), and MediTray products containerized and wrapped.
K080558	December 2008	SteriTite container system and MediTray products containerized and wrapped for STERRAD 200, STERRAD NX, Ozone 125L and pre-vacuum steam flash (IUSS) with disposable filter.
K090068	February 2010	SteriTite container system and MediTray products for AMSCO V-Pro 1 low temperature sterilizer.
K110682	July 2011	SteriTite container system and MediTray products for STERRAD 100NX.
K112904	January 2012	SteriTite container system and MediTray products for AMSCO V-Pro 1 Plus and V-Pro maX sterilizers.
K161415	February 2017	SteriTite container system and MediTray products for STERIZONE VP4 sterilizer.
K173259	June 2018	SteriTite container system and MediTray products for table-top pre-vacuum sterilizers, one (1) year sterility maintenance claim, and five (5) to eight (8) minute dry time claims for terminal steam sterilization.
K212711	April 2022	SteriTite container system and MediTray products for STERIS V-Pro 60 (with lumen claims), V-Pro s2 (with lumen claims), V-Pro maX 2 (with lumen claims), STERRAD 100NX DUO (with lumen claims), Sterizone VP4 lumen claims & Da Vinci Xi Scope and Camera
K221492	October 2022	SteriTite container system and MediTray products compatibility with the STERRAD 100NX Express Cycles & Da Vinci Xi Scope and Camera.
K240898	July 31 2024	Solid bottom SteriTite container compatible with V-Pro maX, V-Pro maX2, V-Pro 60, V-Pro s2.
K252854	June 3 2026	Steam Flush Pressure Pulse (SFPP) compatibility with SteriTite Perforated base and STERRAD 100NX Solid Bottom for DUO and Express Cycles.